
Midmark M11 - v1.0.0
Total Cycles: 2118

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2022 07:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	166.4 F	0.0
2:00	168.7 F	0.0
4:00	191.0 F	0.0
6:00	209.2 F	0.0
8:00	212.1 F	0.0
10:00	212.2 F	0.0
12:00	212.2 F	0.0
14:00	212.2 F	0.0
16:00	212.2 F	0.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 212F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

07/13/2022 07:56 AM