
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000547
10:39 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:07
STERILIZING 4:09
136.0 C 220kPa 4:09
136.2 C 224kPa 4:39
136.1 C 222kPa 5:09
136.4 C 225kPa 5:39
136.2 C 218kPa 6:09
136.4 C 225kPa 6:39
136.1 C 222kPa 7:09
136.1 C 225kPa 7:39
136.1 C 225kPa 7:39
VENTING 7:39
AIR DRYING 8:29
CYCLE COMPLETE 8:33

Digital Signature #
CFDC9AC625DF07E4
