
Midmark M11 - v1.0.5
Total Cycles: 3353

Name: RITTER2
Sterilizer ID: V2313228
02/05/2024 09:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 91.9 F 0.0
4:00 129.8 F 0.3
6:00 130.1 F 2.1
8:00 141.8 F 4.7
10:00 171.3 F 7.7
12:00 207.7 F 11.2
14:00 215.8 F 4.2
16:00 226.4 F 8.1
18:00 234.5 F 12.1
20:00 231.2 F 7.7
22:00 241.9 F 12.6
24:00 239.3 F 10.4
26:00 251.2 F 16.2
28:00 261.6 F 22.2
30:00 271.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.3 F 29.2
2:00 271.5 F 29.3
3:00 271.7 F 28.9
4:00 272.1 F 29.6

Min 271.0 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 224.5 F 0.0
10:00 212.8 F 0.0
15:00 217.5 F 0.0
20:00 216.8 F 0.0
25:00 217.1 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 30:39
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:29
CYCLE COMPLETE

02/05/2024 10:30 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____