

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4385

Name: RITTER2  
Sterilizer ID: V2313228  
02/26/2025 04:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.6 F 0.0  
2:00 173.1 F 3.7  
4:00 199.7 F 9.6  
6:00 220.2 F 6.5  
8:00 231.6 F 11.1  
10:00 233.6 F 9.0  
12:00 237.6 F 9.4  
14:00 248.9 F 15.3  
16:00 262.3 F 22.8  
18:00 272.4 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.2 F 29.3  
2:00 272.5 F 29.7  
3:00 272.0 F 29.7  
4:00 271.9 F 29.2  
  
Min 271.5 F 29.0  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
5:00 220.6 F 0.0  
10:00 187.4 F 0.0  
15:00 182.7 F 0.0  
20:00 180.5 F 0.0  
25:00 180.4 F 0.0  
30:00 180.9 F 0.0  
Drying Complete

FILLING: 4:03  
HEATING: 18:08  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 00:59:35  
CYCLE COMPLETE

02/26/2025 05:32 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_