
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000904
11:34 04/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 3:55
STERILIZING 5:03
136.0 C 227kPa 5:03
136.4 C 232kPa 5:33
136.4 C 232kPa 6:03
136.3 C 231kPa 6:33
136.4 C 233kPa 7:03
136.4 C 232kPa 7:33
136.4 C 232kPa 8:03
136.4 C 232kPa 8:33
136.4 C 232kPa 8:33
VENTING 8:33
AIR DRYING 9:28
CYCLE COMPLETE 24:28

Digital Signature #
F4C1B8966CF12351
