

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5369

Name: RITTER 1  
Sterilizer ID: V2178198  
02/26/2025 01:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.6 F 0.0  
2:00 179.1 F 3.3  
4:00 181.6 F 7.5  
6:00 207.1 F 11.9  
8:00 222.8 F 5.9  
10:00 229.6 F 10.7  
12:00 233.0 F 7.3  
14:00 240.4 F 12.9  
16:00 241.3 F 11.4  
18:00 253.2 F 17.5  
20:00 264.5 F 23.9  
22:00 270.8 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.5 F 29.5  
2:00 271.7 F 29.3  
3:00 272.6 F 29.6  
4:00 271.6 F 29.1

Min 271.6 F 29.0  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 212.0 F 0.0  
10:00 216.3 F 0.0  
15:00 217.6 F 0.0  
20:00 216.2 F 0.0  
25:00 213.7 F 0.0  
30:00 217.6 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 22:46  
STERILIZING: 4:00  
VENTING: 3:58  
DRYING: 30:00  
TOTAL CYCLE: 01:02:28  
CYCLE COMPLETE

02/26/2025 02:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_