
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000803
16:02 17/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:00
PRESSURIZING 5:47
STERILIZING 7:10
136.0 C 220kPa 7:10
136.0 C 219kPa 7:41
136.3 C 222kPa 8:11
136.0 C 221kPa 8:41
136.4 C 222kPa 9:11
135.9 C 219kPa 9:41
136.4 C 223kPa 10:11
136.4 C 221kPa 10:41
136.4 C 221kPa 10:41
VENTING 10:41
AIR DRYING 11:42
CYCLE COMPLETE 11:57

Digital Signature #
14B536C7416E6675
