
Midmark M11 - v1.0.0
Total Cycles: 872

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2021 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.2 F 0.0
2:00 201.9 F 3.4
4:00 206.3 F 8.2
6:00 221.3 F 13.3
8:00 228.8 F 8.5
10:00 240.2 F 12.2
12:00 241.4 F 11.7
14:00 240.6 F 10.5
16:00 253.5 F 17.1
18:00 264.8 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 28.9
2:00 271.8 F 28.3
3:00 271.9 F 28.8
4:00 272.5 F 28.9

Min 271.4 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 221.8 F 0.0
10:00 215.5 F 0.0
15:00 214.1 F 0.0
20:00 227.1 F 0.0
25:00 230.8 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

06/08/2021 11:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____