
Midmark M11 - v1.0.0
Total Cycles: 4043

Name: RITTER 1
Sterilizer ID: V2178198
11/20/2023 03:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.9 F 0.0
2:00 179.6 F 3.0
4:00 190.0 F 7.2
6:00 222.3 F 12.2
8:00 221.4 F 6.3
10:00 234.5 F 11.3
12:00 234.5 F 8.1
14:00 242.4 F 11.5
16:00 245.1 F 12.7
18:00 256.8 F 18.8
20:00 267.4 F 25.4
22:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.6 F 28.5
2:00 271.9 F 28.5
3:00 272.2 F 28.7
4:00 272.5 F 28.9

Min 271.4 F 28.4
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 229.2 F 0.0
10:00 236.4 F 0.0
15:00 229.2 F 0.0
20:00 226.3 F 0.0
25:00 225.9 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

11/20/2023 04:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____