
Midmark M11 - v1.0.5
Total Cycles: 671

Name: RITTER2
Sterilizer ID: V2313228
12/30/2021 11:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.8 F 0.0
2:00 173.0 F 3.6
4:00 197.5 F 9.0
6:00 218.6 F 7.6
8:00 233.1 F 9.8
10:00 234.4 F 8.7
12:00 242.5 F 11.7
14:00 246.6 F 13.8
16:00 259.6 F 21.0
18:00 271.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.0 F 28.7
2:00 272.3 F 29.0
3:00 272.7 F 29.4
4:00 271.8 F 29.2

Min 271.6 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 233.9 F 0.0
10:00 227.3 F 0.0
15:00 212.6 F 0.0
20:00 207.9 F 0.0
25:00 211.4 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 18:35
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

12/30/2021 12:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____