

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1457

Name: RITTER 1  
Sterilizer ID: V2178198  
12/15/2021 09:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.6 F 0.0  
2:00 143.2 F 0.0  
4:00 150.7 F 1.0  
6:00 173.5 F 4.3  
8:00 196.3 F 8.4  
10:00 214.2 F 13.0  
12:00 224.0 F 7.3  
14:00 233.6 F 12.3  
16:00 235.5 F 9.4  
18:00 238.3 F 9.1  
20:00 247.5 F 14.1  
22:00 259.6 F 20.8  
24:00 270.7 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.6 F 29.2  
2:00 271.7 F 28.5  
3:00 271.9 F 28.9  
4:00 272.4 F 29.1

Min 271.4 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.8 F 0.0  
5:00 180.6 F 0.0  
10:00 212.5 F 0.0  
15:00 216.6 F 0.0  
20:00 216.4 F 0.0  
25:00 216.5 F 0.0  
30:00 216.4 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 24:50  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:03:06  
CYCLE COMPLETE

12/15/2021 10:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_