
Midmark M11 - v1.0.0
Total Cycles: 4810

Name: RITTER 1
Sterilizer ID: V2178198
08/13/2024 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.9 F 0.0
2:00 152.2 F 1.2
4:00 175.0 F 4.2
6:00 195.7 F 8.4
8:00 213.0 F 13.5
10:00 224.4 F 6.7
12:00 237.0 F 11.4
14:00 232.6 F 7.8
16:00 244.9 F 12.5
18:00 243.9 F 12.1
20:00 256.3 F 18.6
22:00 267.5 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.9
2:00 271.9 F 28.0
3:00 272.5 F 28.9
4:00 272.5 F 28.6

Min 271.6 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 206.8 F 0.0
10:00 201.6 F 0.0
15:00 198.5 F 0.0
20:00 197.0 F 0.0
25:00 196.8 F 0.0
30:00 195.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:26
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:01:45
CYCLE COMPLETE

08/13/2024 12:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____