

-----  
Midmark M11 - v1.0.5  
Total Cycles: 860

Name: RITTER2  
Sterilizer ID: V2313228  
03/02/2022 03:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.9 F 0.0  
2:00 192.7 F 4.2  
4:00 207.4 F 10.0  
6:00 223.6 F 6.6  
8:00 236.3 F 12.4  
10:00 238.1 F 10.8  
12:00 238.5 F 10.2  
14:00 252.8 F 17.5  
16:00 265.2 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 271.8 F 29.9  
2:00 272.5 F 29.6  
3:00 271.2 F 29.1  
4:00 272.0 F 30.1  
  
Min 271.1 F 29.1  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.7 F 0.0  
5:00 208.5 F 0.0  
10:00 218.0 F 0.0  
15:00 210.6 F 0.0  
20:00 211.8 F 0.0  
25:00 217.0 F 0.0  
30:00 222.4 F 0.0  
Drying Complete

FILLING: 2:29  
HEATING: 17:37  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:16  
CYCLE COMPLETE

03/02/2022 03:58 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_