

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5316

Name: RITTER 1  
Sterilizer ID: V2178198  
02/05/2025 11:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.3 F 0.0  
2:00 187.9 F 2.6  
4:00 192.4 F 6.4  
6:00 203.7 F 11.1  
8:00 216.6 F 4.2  
10:00 229.1 F 9.2  
12:00 238.2 F 10.1  
14:00 238.8 F 10.8  
16:00 237.1 F 8.5  
18:00 250.0 F 15.2  
20:00 261.6 F 21.8  
22:00 272.2 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.6 F 28.7  
2:00 272.1 F 28.7  
3:00 272.3 F 28.5  
4:00 272.3 F 28.9

Min 271.5 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.6 F 0.0  
5:00 203.6 F 0.0  
10:00 220.3 F 0.0  
15:00 230.1 F 0.0  
20:00 232.0 F 0.0  
25:00 224.4 F 0.0  
30:00 225.9 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 22:28  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:01:14  
CYCLE COMPLETE

02/05/2025 12:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_