
Midmark M11 - v1.0.0
Total Cycles: 3822

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2023 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.9 F 0.0
2:00 193.1 F 2.9
4:00 203.5 F 7.7
6:00 224.4 F 13.2
8:00 229.4 F 8.4
10:00 236.8 F 13.5
12:00 239.6 F 10.9
14:00 238.0 F 9.2
16:00 251.4 F 15.8
18:00 262.6 F 22.2
20:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.5
2:00 272.1 F 28.8
3:00 271.8 F 28.2
4:00 272.6 F 28.9

Min 271.3 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 224.3 F 0.0
10:00 232.7 F 0.0
15:00 235.6 F 0.0
20:00 227.0 F 0.0
25:00 225.6 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

09/26/2023 03:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____