

-----  
Midmark M11 - v1.0.5  
Total Cycles: 590

Name: RITTER2  
Sterilizer ID: V2313228  
11/30/2021 03:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
2:00 178.5 F 2.6  
4:00 191.2 F 7.0  
6:00 213.2 F 12.8  
8:00 225.8 F 8.2  
10:00 238.2 F 11.8  
12:00 242.0 F 12.0  
14:00 240.8 F 11.1  
16:00 254.4 F 18.1  
18:00 266.2 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.9  
1:00 272.0 F 28.9  
2:00 272.5 F 29.3  
3:00 272.4 F 29.7  
4:00 271.2 F 29.0  
  
Min 271.2 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
5:00 199.8 F 0.0  
10:00 192.1 F 0.0  
15:00 189.3 F 0.0  
20:00 185.3 F 0.0  
25:00 181.8 F 0.0  
30:00 181.0 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 19:31  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

11/30/2021 04:37 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02