

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001772  
15:35 02/05/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:31  
STERILIZING 3:36  
136.5 C 225kPa 3:36  
136.4 C 226kPa 4:06  
136.4 C 226kPa 4:36  
136.2 C 223kPa 5:06  
136.4 C 224kPa 5:36  
136.4 C 224kPa 6:06  
136.2 C 224kPa 6:36  
136.3 C 225kPa 7:06  
136.3 C 225kPa 7:06  
VENTING 7:06  
AIR DRYING 7:57  
CYCLE COMPLETE 8:29  
-----

Digital Signature #  
8B56D17C30DCE5C1  
-----