

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000821  
14:37 28/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:41  
PRESSURIZING 5:27  
STERILIZING 6:48  
136.0 C 220kPa 6:48  
136.3 C 222kPa 7:18  
136.4 C 221kPa 7:48  
135.9 C 220kPa 8:18  
136.4 C 223kPa 8:48  
136.4 C 221kPa 9:18  
136.3 C 223kPa 9:48  
136.4 C 223kPa 10:18  
136.4 C 223kPa 10:18  
VENTING 10:18  
AIR DRYING 11:19  
CYCLE COMPLETE 41:19

-----  
Digital Signature #  
632A83E5CB0868D8  
-----