
Midmark M11 - v1.0.5
Total Cycles: 3302

Name: RITTER2
Sterilizer ID: V2313228
01/16/2024 09:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 76.2 F 0.0
4:00 89.6 F 0.2
6:00 106.8 F 1.4
8:00 144.7 F 4.1
10:00 170.7 F 7.1
12:00 196.8 F 10.4
14:00 208.4 F 5.0
16:00 217.2 F 7.1
18:00 228.3 F 11.1
20:00 229.1 F 6.9
22:00 241.2 F 12.6
24:00 239.9 F 10.9
26:00 252.4 F 17.1
28:00 263.3 F 23.5
30:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.6
2:00 271.5 F 29.1
3:00 272.6 F 29.6
4:00 271.4 F 29.0

Min 271.4 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 193.5 F 0.0
10:00 197.7 F 0.0
15:00 199.7 F 0.0
20:00 199.7 F 0.0
25:00 199.8 F 0.0
30:00 201.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:17
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:08:22
CYCLE COMPLETE

01/16/2024 10:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____