
Midmark M11 - v1.0.0
Total Cycles: 4158

Name: RITTER 1
Sterilizer ID: V2178198
12/27/2023 12:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.0 F 0.0
2:00 160.0 F 3.1
4:00 170.4 F 6.8
6:00 192.7 F 11.5
8:00 215.7 F 5.3
10:00 227.6 F 10.3
12:00 230.8 F 7.3
14:00 242.3 F 12.4
16:00 240.7 F 10.8
18:00 253.7 F 17.3
20:00 265.0 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.8 F 29.0
2:00 271.3 F 28.3
3:00 272.9 F 29.1
4:00 271.9 F 28.5

Min 271.2 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 199.0 F 0.0
10:00 201.5 F 0.0
15:00 201.0 F 0.0
20:00 199.6 F 0.0
25:00 198.5 F 0.0
30:00 197.0 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

12/27/2023 01:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____