

-----  
Midmark M11 - v1.0.0  
Total Cycles: 587

Name: MIDMARK  
Sterilizer ID: V2178198  
12/29/2020 03:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.8 F 0.0  
2:00 156.8 F 1.9  
4:00 176.6 F 5.7  
6:00 200.2 F 10.8  
8:00 219.5 F 4.2  
10:00 232.9 F 9.8  
12:00 236.3 F 8.2  
14:00 245.7 F 13.1  
16:00 243.6 F 11.8  
18:00 257.2 F 18.9  
20:00 269.1 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.6 F 28.4  
2:00 272.1 F 28.6  
3:00 272.2 F 28.2  
4:00 272.2 F 28.7  
  
Min 271.4 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
5:00 242.4 F 0.0  
10:00 238.0 F 0.0  
15:00 232.9 F 0.0  
20:00 225.3 F 0.0  
25:00 231.9 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 21:02  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:16  
CYCLE COMPLETE

12/29/2020 04:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_