
Midmark M11 - v1.0.0
Total Cycles: 5041

Name: RITTER 1
Sterilizer ID: V2178198
10/31/2024 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.0 F 0.0
2:00 163.1 F 0.1
4:00 160.5 F 0.7
6:00 185.7 F 3.9
8:00 201.2 F 7.6
10:00 211.9 F 11.7
12:00 222.8 F 5.5
14:00 231.3 F 10.2
16:00 235.4 F 8.7
18:00 243.1 F 12.5
20:00 240.6 F 10.7
22:00 252.9 F 16.8
24:00 263.5 F 23.0
26:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 271.3 F 28.4
2:00 272.5 F 28.9
3:00 272.0 F 28.5
4:00 271.2 F 28.1

Min 271.0 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 209.1 F 0.0
10:00 231.3 F 0.0
15:00 236.1 F 0.0
20:00 234.5 F 0.0
25:00 223.5 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:13
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:50
CYCLE COMPLETE

10/31/2024 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____