
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000351
9:20 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:14
STERILIZING 3:17
136.0 C 220kPa 3:17
136.4 C 233kPa 3:47
136.4 C 230kPa 4:17
136.1 C 230kPa 4:47
135.8 C 228kPa 5:17
136.0 C 230kPa 5:47
136.4 C 233kPa 6:17
136.4 C 232kPa 6:47
136.4 C 232kPa 6:47
VENTING 6:47
AIR DRYING 7:38
CYCLE COMPLETE 13:08

Digital Signature #
6F6AF41527276014
