

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1088

Name: RITTER 1  
Sterilizer ID: V2178198  
08/10/2021 02:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.2 F 0.0  
2:00 157.6 F 1.9  
4:00 175.5 F 5.0  
6:00 198.1 F 9.7  
8:00 217.4 F 6.8  
10:00 229.1 F 9.6  
12:00 234.8 F 8.7  
14:00 241.1 F 12.0  
16:00 239.3 F 10.2  
18:00 252.7 F 16.8  
20:00 264.3 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.1  
2:00 271.4 F 28.4  
3:00 272.4 F 29.3  
4:00 272.1 F 28.8  
  
Min 271.2 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.0 F 0.0  
5:00 212.5 F 0.0  
10:00 222.9 F 0.0  
15:00 217.4 F 0.0  
20:00 216.8 F 0.0  
25:00 217.7 F 0.0  
Dry Cycle Aborted

FILLING: 1:21  
HEATING: 21:58  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 27:42  
TOTAL CYCLE: 00:58:08

08/10/2021 03:20 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01