
Midmark M11 - v1.0.5
Total Cycles: 3578

Name: RITTER2
Sterilizer ID: V2313228
05/01/2024 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.2 F 0.0
2:00 148.4 F 2.3
4:00 157.6 F 5.6
6:00 182.8 F 9.5
8:00 202.0 F 9.4
10:00 219.9 F 7.7
12:00 230.1 F 12.3
14:00 231.9 F 9.1
16:00 238.9 F 10.7
18:00 244.4 F 13.4
20:00 256.0 F 19.6
22:00 266.5 F 26.1
24:00 273.0 F 30.6

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 30.2
1:00 272.1 F 30.3
2:00 272.1 F 30.2
3:00 272.2 F 30.1
4:00 272.4 F 30.2

Min 272.0 F 30.1
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 215.9 F 0.0
10:00 194.6 F 0.0
15:00 190.5 F 0.0
20:00 188.3 F 0.0
25:00 185.9 F 0.0
30:00 184.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:22
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:08
CYCLE COMPLETE

05/01/2024 04:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____