
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006989
8:39 27/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:47
STERILIZING 4:48
136.4 C 222kPa 4:48
136.4 C 224kPa 5:18
136.4 C 222kPa 5:48
136.4 C 222kPa 6:18
136.2 C 222kPa 6:48
136.4 C 224kPa 7:18
136.2 C 220kPa 7:48
136.4 C 221kPa 8:18
136.4 C 221kPa 8:18
VENTING 8:18
AIR DRYING 9:13
CYCLE COMPLETE 9:16

Digital Signature #
2AEB0851FC5E7A97
