
Midmark M11 - v1.0.0
Total Cycles: 4047

Name: RITTER 1
Sterilizer ID: V2178198
11/21/2023 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.7 F 0.0
2:00 161.1 F 2.5
4:00 174.0 F 5.9
6:00 204.4 F 9.9
8:00 223.1 F 5.7
10:00 227.1 F 8.3
12:00 237.9 F 13.3
14:00 239.2 F 10.5
16:00 236.8 F 8.6
18:00 249.9 F 15.0
20:00 261.0 F 21.3
22:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 28.9
2:00 272.3 F 29.1
3:00 272.2 F 28.8
4:00 272.5 F 29.2

Min 271.5 F 28.5
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 234.9 F 0.0
10:00 223.5 F 0.0
15:00 229.6 F 0.0
20:00 230.2 F 0.0
25:00 230.4 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:02:50
CYCLE COMPLETE

11/21/2023 12:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____