
Midmark M11 - v1.0.0
Total Cycles: 2771

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2023 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 82.7 F 0.0
4:00 98.5 F 0.0
6:00 135.1 F 0.0
8:00 152.8 F 0.0
10:00 185.7 F 1.4
12:00 200.3 F 4.9
14:00 212.0 F 8.0
16:00 222.8 F 11.5
18:00 225.4 F 5.8
20:00 236.9 F 10.6
22:00 234.5 F 7.9
24:00 245.8 F 13.4
26:00 243.5 F 11.8
28:00 255.7 F 18.2
30:00 266.7 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.6
2:00 271.6 F 28.2
3:00 272.3 F 28.6
4:00 271.5 F 28.1

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 195.8 F 0.0
10:00 214.5 F 0.0
15:00 225.0 F 0.0
20:00 224.1 F 0.0
25:00 219.8 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:33
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:31
CYCLE COMPLETE

01/12/2023 09:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____