
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008095
10:34 22/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:20
STERILIZING 4:30
136.0 C 220kPa 4:30
135.9 C 221kPa 5:00
135.8 C 219kPa 5:30
136.1 C 223kPa 6:00
136.2 C 222kPa 6:30
136.3 C 222kPa 7:00
136.4 C 221kPa 7:30
136.3 C 222kPa 8:00
136.3 C 222kPa 8:00
VENTING 8:00
AIR DRYING 8:50
CYCLE COMPLETE 38:50

Digital Signature #
6E8454EA8F52F347
