
Midmark M11 - v1.0.5
Total Cycles: 4493

Name: RITTER2
Sterilizer ID: V2313228
04/07/2025 01:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.8 F 0.0
2:00 107.2 F 0.7
4:00 122.9 F 2.8
6:00 149.5 F 5.8
8:00 177.8 F 9.3
10:00 195.8 F 13.2
12:00 216.6 F 5.3
14:00 225.3 F 9.1
16:00 235.8 F 13.3
18:00 235.5 F 9.5
20:00 239.9 F 10.5
22:00 244.3 F 13.0
24:00 255.5 F 18.9
26:00 265.6 F 25.1
28:00 272.1 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.7
1:00 273.0 F 30.3
2:00 271.5 F 29.8
3:00 273.2 F 30.0
4:00 271.5 F 30.0

Min 271.3 F 29.6
Max 273.4 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 228.1 F 0.0
10:00 217.6 F 0.0
15:00 181.2 F 0.0
20:00 178.7 F 0.0
25:00 180.0 F 0.0
30:00 177.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:22
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:07:15
CYCLE COMPLETE

04/07/2025 03:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____