
Midmark M11 - v1.0.5
Total Cycles: 2807

Name: RITTER2
Sterilizer ID: V2313228
08/28/2023 04:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.7 F 0.0
2:00 173.1 F 0.2
4:00 195.3 F 4.2
6:00 216.6 F 8.6
8:00 230.5 F 12.7
10:00 232.5 F 9.8
12:00 236.2 F 9.2
14:00 245.0 F 13.2
16:00 244.7 F 12.9
18:00 257.1 F 19.5
20:00 268.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.3
2:00 271.4 F 29.0
3:00 272.3 F 29.2
4:00 271.3 F 28.7

Min 271.3 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 208.5 F 0.0
10:00 210.5 F 0.0
15:00 214.1 F 0.0
20:00 211.4 F 0.0
25:00 211.3 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

08/28/2023 05:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____