

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1077

Name: RITTER2  
Sterilizer ID: V2313228  
05/05/2022 10:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.2 F 0.0  
2:00 145.5 F 2.4  
4:00 172.1 F 6.5  
6:00 198.6 F 11.5  
8:00 217.0 F 5.8  
10:00 230.4 F 11.2  
12:00 232.0 F 8.4  
14:00 240.3 F 11.0  
16:00 245.5 F 13.5  
18:00 257.9 F 20.1  
20:00 268.7 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.4 F 29.4  
2:00 271.7 F 29.3  
3:00 272.1 F 29.3  
4:00 271.8 F 29.5

Min 271.3 F 28.8  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
5:00 212.3 F 0.0  
10:00 186.3 F 0.0  
15:00 174.4 F 0.0  
20:00 165.0 F 0.0  
25:00 166.4 F 0.0  
30:00 171.4 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 21:08  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:56  
CYCLE COMPLETE

05/05/2022 11:28 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02