
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002358
11:50 05/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:02
STERILIZING 3:07
136.0 C 221kPa 3:07
136.2 C 224kPa 3:37
136.3 C 224kPa 4:07
136.2 C 224kPa 4:37
136.4 C 225kPa 5:07
136.4 C 225kPa 5:37
136.3 C 225kPa 6:07
136.4 C 226kPa 6:37
136.4 C 226kPa 6:37
VENTING 6:37
AIR DRYING 7:30
CYCLE COMPLETE 7:34

Digital Signature #
459013267983FDC3
