

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002144  
9:41 15/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:19  
PRESSURIZING 2:25  
STERILIZING 3:28  
136.0 C 221kPa 3:28  
136.4 C 222kPa 3:58  
136.4 C 222kPa 4:28  
136.4 C 223kPa 4:58  
136.2 C 223kPa 5:28  
136.4 C 224kPa 5:58  
136.3 C 225kPa 6:28  
136.2 C 222kPa 6:58  
136.2 C 222kPa 6:58  
VENTING 6:58  
AIR DRYING 7:49  
CYCLE COMPLETE 37:49  
-----

Digital Signature #  
DAA94F7502390774  
-----