

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.9uS / 0.5ppm  
CYCLE NUMBER 000305  
8:16 20/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:50  
PRESSURIZING 3:52  
STERILIZING 5:04  
136.0 C 229kPa 5:04  
136.4 C 232kPa 5:34  
136.1 C 231kPa 6:04  
136.2 C 231kPa 6:34  
136.4 C 231kPa 7:04  
136.4 C 233kPa 7:34  
CYCLE ABORTED 8:03  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 8:03  
-----

Digital Signature #  
36EDF6A06AA3D199  
-----