
Midmark M11 - v1.0.0
Total Cycles: 1353

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2021 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 190.6 F 3.3
4:00 211.0 F 7.4
6:00 223.8 F 12.9
8:00 228.8 F 7.7
10:00 239.7 F 13.0
12:00 238.4 F 10.4
14:00 237.6 F 8.4
16:00 250.2 F 15.4
18:00 261.9 F 22.1
20:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.8
2:00 271.8 F 28.7
3:00 272.3 F 28.7
4:00 271.7 F 28.6

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 225.3 F 0.0
10:00 236.6 F 0.0
15:00 236.7 F 0.0
20:00 226.3 F 0.0
25:00 220.6 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

11/10/2021 01:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____