

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007037  
10:59 06/03/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 2:58  
STERILIZING 3:56  
136.0 C 220kPa 3:56  
136.0 C 218kPa 4:26  
136.2 C 223kPa 4:56  
136.3 C 223kPa 5:26  
136.4 C 223kPa 5:56  
136.4 C 224kPa 6:26  
136.1 C 222kPa 6:56  
136.2 C 222kPa 7:26  
136.2 C 222kPa 7:26  
VENTING 7:26  
AIR DRYING 8:14  
CYCLE COMPLETE 8:47

-----  
Digital Signature #  
00E35118E7A5BCF8  
-----