
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 000837
14:21 16/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:37
STERILIZING 4:36
136.0 C 220kPa 4:36
136.4 C 224kPa 5:06
136.4 C 224kPa 5:36
136.3 C 224kPa 6:06
136.5 C 225kPa 6:36
136.4 C 224kPa 7:06
136.3 C 223kPa 7:36
136.4 C 224kPa 8:06
136.4 C 224kPa 8:06
VENTING 8:06
AIR DRYING 9:01
CYCLE COMPLETE 9:06

Digital Signature #
016B7415B26AA802
