
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001710
14:03 07/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:50
STERILIZING 3:56
136.4 C 223kPa 3:56
136.3 C 224kPa 4:26
136.4 C 225kPa 4:56
136.4 C 224kPa 5:26
136.3 C 225kPa 5:56
136.3 C 224kPa 6:26
136.4 C 224kPa 6:56
136.3 C 223kPa 7:26
136.3 C 223kPa 7:26
VENTING 7:26
AIR DRYING 8:16
CYCLE COMPLETE 8:22

Digital Signature #
9B90C0780FED8DF1
