

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1892

Name: RITTER 1  
Sterilizer ID: V2178198  
05/02/2022 10:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
2:00 195.1 F 4.3  
4:00 204.3 F 9.2  
6:00 219.2 F 7.5  
8:00 230.0 F 9.6  
10:00 235.8 F 8.7  
12:00 243.7 F 13.0  
14:00 242.7 F 11.7  
16:00 255.9 F 18.6  
18:00 267.4 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.6 F 29.3  
2:00 271.9 F 28.6  
3:00 271.7 F 28.9  
4:00 272.6 F 29.2  
  
Min 271.3 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 223.8 F 0.0  
10:00 229.9 F 0.0  
15:00 226.7 F 0.0  
20:00 231.1 F 0.0  
25:00 229.1 F 0.0  
30:00 228.5 F 0.0  
  
Drying Complete

FILLING: 2:28  
HEATING: 19:26  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:03  
CYCLE COMPLETE

05/02/2022 11:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_