
Midmark M11 - v1.0.0
Total Cycles: 1100

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2021 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.8 F 0.0
2:00 196.8 F 4.4
4:00 216.3 F 10.3
6:00 225.5 F 5.9
8:00 240.5 F 12.5
10:00 239.9 F 10.5
12:00 237.7 F 9.2
14:00 252.8 F 16.5
16:00 264.9 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 272.7 F 28.7
2:00 271.2 F 28.1
3:00 272.3 F 28.5
4:00 271.2 F 28.1

Min 271.2 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 236.6 F 0.0
10:00 224.9 F 0.0
15:00 228.4 F 0.0
20:00 219.1 F 0.0
Dry Cycle Aborted

FILLING: 1:59
HEATING: 17:38
STERILIZING: 4:00
VENTING: 3:06
DRYING: 21:01
TOTAL CYCLE: 00:47:44

08/19/2021 12:35 PM
ACCEPTED: X
REJECTED:
Operator: U02