
Midmark M11 - v1.0.0
Total Cycles: 4484

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2024 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 161.0 F 2.0
4:00 195.6 F 5.1
6:00 200.2 F 9.1
8:00 214.1 F 13.4
10:00 220.7 F 6.6
12:00 231.0 F 10.9
14:00 231.4 F 7.1
16:00 243.0 F 12.4
18:00 240.4 F 10.5
20:00 252.3 F 16.4
22:00 262.8 F 22.5
24:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.9 F 29.3
2:00 271.8 F 28.6
3:00 272.2 F 29.2
4:00 272.7 F 29.1

Min 271.4 F 28.5
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 235.8 F 0.0
10:00 226.6 F 0.0
15:00 234.5 F 0.0
20:00 227.9 F 0.0
25:00 228.5 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 25:17
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:16
CYCLE COMPLETE

04/24/2024 10:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____