
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000390
6:48 02/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:08
STERILIZING 4:21
136.5 C 224kPa 4:21
136.2 C 220kPa 4:51
136.2 C 220kPa 5:21
136.4 C 223kPa 5:51
136.4 C 223kPa 6:21
136.4 C 220kPa 6:51
136.3 C 222kPa 7:21
136.1 C 221kPa 7:51
136.1 C 221kPa 7:51
VENTING 7:52
AIR DRYING 8:42
CYCLE COMPLETE 8:52

Digital Signature #
4C163D65539769F1
