

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1956

Name: RITTER2  
Sterilizer ID: V2313228  
01/23/2023 10:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.2 F 0.0  
2:00 146.0 F 0.1  
4:00 159.6 F 0.5  
6:00 184.9 F 4.5  
8:00 210.1 F 9.6  
10:00 223.2 F 7.2  
12:00 232.4 F 11.4  
14:00 235.0 F 9.5  
16:00 237.9 F 9.2  
18:00 249.3 F 15.5  
20:00 261.3 F 22.3  
22:00 272.4 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.6 F 30.1  
2:00 272.0 F 29.4  
3:00 272.8 F 29.7  
4:00 272.4 F 30.0

Min 270.9 F 28.7  
Max 273.3 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
5:00 230.6 F 0.0  
10:00 213.1 F 0.0  
15:00 208.1 F 0.0  
20:00 209.6 F 0.0  
25:00 224.6 F 0.0  
30:00 234.1 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 22:23  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:37  
CYCLE COMPLETE

01/23/2023 11:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_