
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004303
12:57 06/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:08
PRESSURIZING 1:15
STERILIZING 2:20
136.4 C 220kPa 2:20
136.3 C 224kPa 2:50
136.4 C 223kPa 3:20
136.4 C 223kPa 3:50
136.4 C 223kPa 4:20
136.3 C 224kPa 4:50
136.2 C 223kPa 5:20
136.4 C 223kPa 5:50
136.4 C 223kPa 5:50
VENTING 5:50
AIR DRYING 6:39
CYCLE COMPLETE 6:48

Digital Signature #
0D9B92B81D202BE9
