
Midmark M11 - v1.0.5
Total Cycles: 577

Name: RITTER2
Sterilizer ID: V2313228
11/25/2021 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 73.7 F 0.0
4:00 83.7 F 0.0
6:00 121.6 F 0.1
8:00 156.2 F 0.1
10:00 187.6 F 0.3
12:00 205.3 F 1.6
14:00 215.4 F 4.4
16:00 223.3 F 7.3
18:00 233.7 F 10.7
20:00 232.2 F 8.8
22:00 239.5 F 11.3
24:00 236.2 F 9.3
26:00 239.5 F 10.1
28:00 248.3 F 14.6
30:00 259.2 F 20.7
32:00 269.3 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.3
2:00 271.6 F 28.5
3:00 272.3 F 29.4
4:00 272.2 F 28.9

Min 271.3 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 220.6 F 0.0
10:00 193.7 F 0.0
15:00 191.3 F 0.0
20:00 195.9 F 0.0
25:00 194.2 F 0.0
30:00 190.7 F 0.0

Drying Complete

FILLING: 1:14
HEATING: 33:06
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:11:39
CYCLE COMPLETE

11/25/2021 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____