
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000031
11:52 09/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:47
PRESSURIZING 0:53
STERILIZING 2:06
136.3 C 222kPa 2:06
136.4 C 222kPa 2:36
136.4 C 224kPa 3:06
136.3 C 225kPa 3:36
136.4 C 224kPa 4:06
136.4 C 225kPa 4:36
136.4 C 223kPa 5:06
136.4 C 225kPa 5:36
136.4 C 225kPa 5:36
VENTING 5:36
AIR DRYING 6:36
CYCLE COMPLETE 6:43

Digital Signature #
B9F33966BC78B3F1
