
Midmark M11 - v1.0.0
Total Cycles: 1744

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2022 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 73.9 F 0.0
4:00 82.6 F 0.1
6:00 120.4 F 0.7
8:00 156.0 F 2.6
10:00 172.8 F 5.1
12:00 195.2 F 8.8
14:00 214.3 F 13.1
16:00 221.3 F 5.8
18:00 232.6 F 9.9
20:00 238.3 F 9.8
22:00 240.5 F 11.1
24:00 236.1 F 8.7
26:00 250.1 F 15.3
28:00 261.4 F 21.6
30:00 271.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.5 F 29.0
2:00 271.1 F 28.2
3:00 272.4 F 29.1
4:00 271.9 F 28.4

Min 271.0 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 192.9 F 0.0
10:00 201.9 F 0.0
15:00 195.1 F 0.0
20:00 199.6 F 0.0
25:00 217.8 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:36
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

03/25/2022 08:54 AM
ACCEPTED: X
REJECTED:
Operator: U02