
Midmark M11 - v1.0.5
Total Cycles: 885

Name: RITTER2
Sterilizer ID: V2313228
03/16/2022 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.6 F 0.0
2:00 183.9 F 3.7
4:00 198.5 F 9.0
6:00 222.6 F 8.1
8:00 233.1 F 10.9
10:00 234.7 F 8.9
12:00 238.3 F 9.7
14:00 249.2 F 15.2
16:00 261.7 F 22.4
18:00 273.0 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.5 F 30.0
2:00 272.6 F 29.4
3:00 272.7 F 30.2
4:00 272.4 F 29.5

Min 271.0 F 28.8
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 250.3 F 0.0
10:00 233.6 F 0.0
15:00 223.6 F 0.0
20:00 219.7 F 0.0
25:00 221.2 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 2:24
HEATING: 18:14
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:46
CYCLE COMPLETE

03/16/2022 02:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____