
Midmark M11 - v1.0.5
Total Cycles: 3349

Name: RITTER2
Sterilizer ID: V2313228
02/02/2024 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.5 F 0.0
2:00 133.3 F 0.5
4:00 142.5 F 2.4
6:00 172.6 F 6.0
8:00 196.8 F 10.2
10:00 214.5 F 6.1
12:00 224.5 F 8.7
14:00 236.6 F 12.4
16:00 238.4 F 11.2
18:00 237.9 F 9.8
20:00 251.2 F 16.3
22:00 262.5 F 23.0
24:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 271.7 F 29.2
2:00 271.9 F 29.5
3:00 272.2 F 29.7
4:00 271.7 F 29.6

Min 271.4 F 29.1
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 222.4 F 0.0
10:00 189.8 F 0.0
15:00 181.7 F 0.0
20:00 177.6 F 0.0
25:00 175.1 F 0.0
30:00 173.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:57
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:03:20
CYCLE COMPLETE

02/02/2024 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____